


3D fluorescence microscopy

For 3D, acquire a “focal series” (stack) of images:


Take an image, refocus the sample, take another image, refocus, etc.

Problem: Each image contains out-of-focus blur from other focal planes

Approach 1:
Physically exclude the blur by
confocal microscopy




Approach 2:
Remove the blur computationally



Deconvolution of 3D data


(Dividing fission yeast cell)

Raw data



4 μm

Deconvolved



4 μm


The Fourier Transform

$$\tilde{g}(k) = \int g(x) e^{-2\pi i k x} dx$$

Any (nice) function $g(x)$ can be *equally well* described as a sum of waves.


The “Fourier transform” $\tilde{g}(k)$ specifies the amplitude A and the phase ϕ for the component wave of wavelength $L = 1/k$

Long wavelength (low resolution) info is close to the origin




Fourier Transform and Digital Filter Applets

<http://www.falstad.com/mathphysics.html>




Low Resolution Duck


A Duck



and its scattering function (Fourier Transform):




a blurred duck



$\Leftarrow$

Using only low resolution data



Convolutions

$$(f \otimes g)(r) = \int f(a) g(r-a) da$$

Why do we care?

- They are everywhere...

- The **convolution theorem**:

If $h(r) = (f \otimes g)(r),$

then $\hat{h}(k) = \hat{f}(k) \hat{g}(k)$


A convolution in real space becomes a product in reciprocal space & vice versa

Symmetry: $g \otimes f = f \otimes g$


So what is a convolution, intuitively?

- “Blurring”

- “Drag and stamp”



2D PSF, OTF of an in-focus lens viewing a point



α = maximum angle

$N.A. = n \sin \alpha$,

the highest spatial frequency is then $f_c = (2n \sin \alpha)/\lambda$
 $f_c = 0.178 \mu\text{m}^{-1}$ for $\lambda = 500\text{nm}$, 1.4NA oil immersion lens

using Raleigh's criterion, the smallest separation
 between two points that can be resolved is $1/(1.22 \cdot f_c)$
 $d_{min} = 0.146 \mu\text{m}$




Figure 3 The in-focus contrast-transfer functions for several Zeiss lenses are compared. They are (1) 10×0.25 N.A., (2) 25×0.5 N.A., (3) 50×1.0 N.A. oil, (4) 100×1.25 N.A. oil, (5) 63×1.4 N.A. oil.








Figure 4 The point-spread functions calculated by Fourier transformation of the CTFs of Figure 3. (A) 25×0.5 N.A. lens (#2); (B) 63×1.4 N.A. lens (#5). Each grid square is $\lambda/6$ or $833 \text{ \AA}$. Note how much sharper the 1.4 N.A. lens is than the 0.5 N.A. lens.


OTF, PSF of a defocused lens



25x lens 0 μm defocus




25x lens 2.5 μm defocus





25x lens 5 μm defocus


3D OTF




Only a bowl-shaped segment of the shell makes it through the objective:
The intensity now becomes a convolution of the **bowl** with itself, which has
a donut-shaped region of support.




Experimentally measured OTF



- Intensity is very peaked at origin
- data along Z direction is missing




Experimentally measured PSF



PSF, OTF & deconvolution

In real space:



In reciprocal space the convolution becomes a product:

$$\widetilde{\text{Image}} = \widetilde{\text{Object}} \cdot \widetilde{\text{PSF}}$$

This suggests:


$$\widetilde{\text{Object}} = \frac{\widetilde{\text{Image}}}{\text{OTF}}$$

??? ("Deconvolution")

PSF is called the
"Optical Transfer Function",
OTF

PSF, OTF & deconvolution

In real space:



In reciprocal space the convolution becomes a product:

$$\widehat{\text{Image}} = \widehat{\text{Object}} \cdot \widehat{\text{PSF}}$$

This suggests:

$$\widehat{\text{Object}} = \frac{\widehat{\text{Image}}}{\widehat{\text{OTF}}}$$

??? ("Deconvolution")

PSF is called the
"Optical Transfer Function",
OTF

What's the catch??

A: We can't divide by OTF(k) if it is **zero** (or small because of noise)

Deconvolution strategies

Nearest neighbor: simplest method, only takes into account adjacent sections
ie subtract out blurred version of adjacent sections from central section


$$I_j \approx c_1[O_j - c_2(\text{PSF}_{\Delta z} \otimes O_{j+1} + \text{PSF}_{-\Delta z} \otimes O_{j-1})]$$

for speed do the convolutions as Fast Fourier Transforms (FFT), multiplication, FFT^{-1}

$$\text{PSF}_{\Delta z} \otimes O_{j+1} = \text{FFT}^{-1}[\text{FFT}(O_{j+1}) \cdot \text{OTF}_{\Delta z}]$$

optical section from a DAPI
stained polytene nucleus

before and after nearest neighbor



Deconvolution strategies

much better to consider the contributions of all the sections to one another = 3D

Weiner filter: simplest 3D method, linear processing, takes care of zeros

$$\widetilde{\text{Object}} = \frac{\widetilde{\text{Image}}}{\text{OTF} + \gamma}$$

γ is related to the signal to noise, sets maximum amplification

or if OTF is complex:


$$\frac{\widetilde{\text{Image}} \cdot \text{OTF}^*}{\text{OTF} \cdot \text{OTF}^* + \gamma}$$

Deconvolution strategies

even better to include *a priori* knowledge about solution such as positivity: $\text{Object} \geq 0$

family of iterative constrained methods

work by calculating convolution, followed by update



arithmetic update (vanCittert's method)

start with $I^0 = O$

$$I^{k+1} = I^k + (O - \text{PSF} \otimes I^k) \quad \text{with positivity constraint: if } I^{k+1} < 0 \text{ then set } I^{k+1} = 0$$


or use multiplicative update (Gold's method)

$$I^{k+1} = I^k \cdot \frac{O}{\text{PSF} \otimes I^k}$$


Deconvolution strategies

what if your PSF is not accurate?

blind deconvolution seeks to estimate both Image and PSF



Some comparisons: Hela cells, DAPI stained




now many related variations


practical issues are signal to noise, accuracy of PSF

if sample is very thick then OSF can vary throughout sample
(minimize by matching index of refraction)


Properties of the Sample Affect Image Formation in Microscopy




Properties of the Sample Affect Image Formation in Microscopy



Properties of the Sample Affect Image Formation in Microscopy





Depth-Dependent Effects in Water: Simulations of PSFs



Depth Dependent Deconvolution: Biological Data

Line Profile Through 1 Spot in Nucleus

Section $\sim 7.6 \mu\text{m}$ into sample



Hanser

Adaptive optics can correct for depth dependent effects (reshape optical wavefront)





Fig. 2. Microscope layout. Grey represents the emission path. The blue striped beam is the excitation light and the red striped beam is the reference beam. See text for details.

Adaptive optics can correct for depth dependent effects


flat mirror

mirror shaped to correct
spherical aberration



raw

decon



Alexa 488-phalloidin stained mouse cells:
images at $4.4\mu\text{m}$ below cover slip

Structured illumination microscopy: the idea


Two patterns
superposed
multiplicatively
give rise to
moiré fringes



The moiré fringes
may be coarse enough
to resolve
even if neither
original pattern is

- *Illuminate the sample with a light pattern*
- *Observe moiré fringes between the pattern and the sample structure*
- *Deduce otherwise unresolvable information about the sample*


Illuminate sample with parallel stripes



*The emitted light contains 3 superimposed
information components shifted by $0, \pm$ the inverse stripe spacing*

Record 3 images ($0^\circ, 120^\circ, 240^\circ$ shifts) to sort out 3 components


Resolution extension by Structured Illumination



Q: What about 3D?

A: Use 3 beams


(admit diffraction orders 1, 0 and +1)



3 coherent
beams





7 intensity
components




Covers full "confocal" OTF support
without discarding light

Same procedure as for 2D,
except need 5 phases


illumination intensity in real space




Observable Region
Structured Illumination (3 orientations)




Structured illumination apparatus






Reconstruction in reciprocal space




Resolution comparison
Actin in a HeLa cell



Conventional microscopy



**Structured illumination
final reconstruction**



Actin in a HeLa cell




170
nm




120 nm microspheres

Conventional
microscopy



Structured illumination
microscopy





120 nm microspheres



Conventional microscopy

Structured illumination microscopy


XY





XZ




350 nm
axial
FWHM








What are the limits of Structured Illumination?




- Linear theory says this is impossible
- How about exploiting non-linear processes?

A simple source of nonlinearity: saturation


(R. Heintzmann Max-Planck Gottingen)



Saturated Structured Illumination




FT (log scale)



Resolution extension by nonlinear structured illumination


Effective observable regions

Conventional
microscopy




200 nm res.

Linear
Structured
illumination
3 directions



100 nm res.


Nonlinear structured illumination
2 new harmonics, 8 directions



50 nm res.


Non-linear structured illumination

Conventional
microscopy




**~250 nm resolution
(diffraction limit)**

Linear
structured
illumination



~120 nm resolution

Saturated
structured
illumination




46 nm resolution!

50 nm microspheres


Drosophila embryo section DNA/RNA stain

Conventional
microscopy




Min. FWHM $\approx$ 280 nm

Linear
structured
illumination



Min. FWHM $\approx$ 110 nm

Saturated
structured
illumination
(1 new harmonic)



Min. FWHM $\approx$ 80 nm

Going beyond the diffraction limit: more light collection angles

We know:


Higher NA



Gathering light over
larger set of angles




Higher resolution




...So what about gathering the light
emitted toward the back side?

OTF when detecting through two lenses




Free space

support of
 $\tilde{A}(\mathbf{k})$




Convolved
with itself:


support of
 $\text{OTF}_{\text{det}}(\mathbf{k})$




Through one
objective




Convolved
with itself:




Through two
objectives




Convolved
with itself:




I⁵M concept




I⁵M OTF




Detection angles




Illum. angles




OTF_{det}



illum



OTF_{eff} =




OTF_{det}

⊗

⊗


illum

=




OTF_{eff}

Observable Information [Optical Transfer Functions]




Deconvolution of bead data


One lens




Two lens det. & illum.



After Wiener filtering




After iterative
deconvolution
using
positivity
constraint



FWHM
~70nm

Microtubules in a recently divided PtK2 cell


XY
view
of area




conventional microscope
+ deconvolution




XZ
projections
1 μ m slices




15M
3D view




Combine methods: I⁵S
use two lenses to collect more angles



Observable Region through 2 opposing objectives
Structured Illumination (3 orientations) – I⁵S




Comparison of Resolving Power



Sample: $0.12\mu\text{m}$ red-fluorescence microsheres

Lin Shao, Mats Gustafsson

Drosophila Anaphase Chromosomes (.2 μ m wide)



Lin Proc.


Decon

Decon + LCE

Lin Shao

Comparison between OM and EM-Tomo

EM



OM

